CE

SPORTON LAB.

NVLAP LAB CODE 200079-0

Certificate No: EC661506

CERTIFICATE

- **EQUIPMENT : Adaptor**

MODEL NO. : ASLX-Y Series (Single output), ASLXD-Z Series (Dual outputs) ASMX-Y Series Single output), ASMXD-Z Series (Dual outputs) where X=40 or 60 used to represetn total output power of 40 or 60 Watts, where Y=5, 6, 7.5, 9, 10, 12, 13.5, 15, 18, 20, 24, 27, 30, 36, 48 used to represent single output voltage (Vdc), where D=dual output voltage, and where Z=1, 2, 3, 5

APPLICANT : ASTRODYNE CORORATION
300 Myles Standish Boulevard Taunton, MA 02780 USA

SPORTON LAB.

I HEREBY CERTIFY THAT:

THE MEASUREMENTS SHOWN IN THIS TEST REPORT WERE MADE IN ACCORDANCE WITH THE PROCEDURES GIVEN IN **EUROPEAN COUNCIL DIRECTIVE 89/336/EEC.** THE EQUIPMENT WAS **PASSED** THE TEST PERFORMED ACCORDING TO **European Standard EN 60601-1-2:2001 (EN55011:1998/A1:1999/A2:2002, EN 61000-3-2:2000, EN 61000-3-3:1995/A1:2001, IEC 61000-4-2:1995/A2:2000, IEC 61000-4-3:2002, IEC 61000-4-4:1995/A2:2001, IEC 61000-4-5:1995/A1:2000, IEC 61000-4-6:1996/A1:2000, IEC 61000-4-8:1993/A1:2000, IEC 61000-4-11:1994/A1:2000).** THE TEST WAS CARRIED OUT ON **Jul. 1, 2006** AT **SPORTON INTERNATIONAL INC.** LAB.

July 04, 2006

Alex Chen
Manager